



EUROPEAN UNION

DELEGATION OF THE EUROPEAN COMMISSION
TO THE REPUBLIC OF SERBIA

Belgrade, 10/09/2009

Ref. (D) 03666 VMA/SND

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 3

Supply of equipment and vehicles for the treatment of the healthcare waste Serbia

Publication ref.: EuropeAid/128417/C/SUP/RS

Tender no: 08SER01/31/21

No	Question	Answer
1.	<p>With reference to the Delivery/Distribution List enclosed to the Clarifications no2 issued by the Contracting Authority, we kindly ask you to confirm that:</p> <p>A. For LOT n°2 the Delivery Address is:</p> <p>Ministry of health Omladinskih brigade 1 11000 Belgrade, Serbia</p> <p>As specified in Annex II – <u>Technical Specifications</u>. Please kindly confirm.</p> <p>B. Consequently we understand that the “Recommended Distribution List” attached to the Contracting Authority Clarifications no° list is all the Beneficiaries of the vehicles of LOT n°2 and we assume that the transportation from the above-mentioned Delivery Address to the different beneficiaries will be borne totally by the Contracting</p>	<p>The correct delivery address for Lot 2 is:</p> <p>Ministry of Health, Nemanjina 22-24, 11000 Belgrade</p> <p>Please also see the correction of the recommended distribution list attached, for your information.</p>

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 3

	<p>Authority. Please kindly confirm.</p>	
2.	<p>Finally the Total of the Small Van (i.e. 41) and the Large Van (i.e. 10), as specified in the above-mentioned "Recommended Distribution List", doesn't correspond to the quantity indicated in the Annex II – Technical Specifications, that is Small Van n°47 and the Large Van n°7. Please kindly specify the exact quantity.</p>	<p>The Annex II and Distribution List should read: 47 small vans and 7 large vans, as cited in the Technical specifications.</p>
3.	<p>Container washing set: Container is made from Polyethylene (HD-PE!) (Advantage: corrosion free, light weight; 1.7 kg container only!) Please put an option to be able to offer Polyethylene container. No manometer (but a red marking on bleeder valve will indicate the max. pressure. If pressure is too low, then the operator will realise that by having a poor spray volume and this will indicate, to pump up the sprayer again.) Please confirm if this specification is acceptable.</p>	<p>Please follow the requirements of the technical specifications from tender dossier.</p>
4.	<p>Container washing set: No certificates are available, since the sprayer does not fall under the European Regulation for pressure devices # 97/23/EG. <u>So please omit this requirement from the specifications.</u></p>	<p>The UN standard is available on line. The certificate is issued by the manufacturer of the container and in most cases the container itself is stamped to verify compliance with UN 6.2. CE and GS certificates are not acceptable. The UN certificate is the only certificate that is acceptable.</p>
5.	<p>Container washing set: Different nozzles are not needed since a vario nozzle, with adjustable spray jet is included. <u>This will allow to spray a round spray jet or a cone spray jet and is individually adjustable! Please change the requirement accordingly.</u></p>	<p>Please follow the requirements of the technical specifications from tender dossier.</p>
6.	<p>For Item 8 of Lot-1, Logistic Set,</p>	<p>Yes a set is supplied with each steam based treatment system.</p>

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 3

	<p>there are twenty (20) transport containers, 20.000 plastic bags, 1000 and 1500 pieces sharp containers. Does this set will be supplied with each steam based infectious waste treatment systems or do they total numbers for Lot-1.</p>	
--	--	--