



Belgrade, 11/10/2011

DELSRBEG-FCT/2011/D/03465

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 1

Project title: Supply of equipment for the National Referent Laboratories Directorate - Re-launch 2

Publication ref.: EuropeAid/130476/C/SUP/RS

Tender no.: 10SER01/22/13

| No. | Question | Answer |
|-----|---|--|
| 1. | <p>Lot 2 Item 76 In the technical specifications the following requirement is mentioned: "Suitable for at least 60-90 mm Petri Dishes (or more)". We herewith kindly would like to ask if a range of 55-90mm Petri Dishes is acceptable to you?</p> | <p>Technical Specifications are the minimum required.</p> |
| 2. | <p>Instructions to Tenderers article 14.1. We refer to the site visit which will be held on 14th October 2011 at 10:00 CET. As we are planning to attend this visit we herewith kindly request you to send us timely before the site visit accurate technical drawings of the building. Also if possible please provide us with accurate information of where each item of the supplied equipment will be placed within the building. Having this information in combination with the site visit will allow us to present you with an accurate offer for this project regarding installation,</p> <p>We thank you in advance for your kind co-operation and look forward to your soonest reply. Taking into account the site visit will be held on 14th October 2011 at 10:00 CET we would appreciate if drawings could be send by 10th October 2010.</p> | <p>In line with the policy of the Beneficiary, the technical drawings of the premises are not in the public domain. The outline of the premises may be shared with the selected Contractor(s) upon their request.</p> <p>Please also be reminded to read carefully the general provisions of the Technical Specifications, which, among others indicated the following:</p> <p>"The beneficiary premises are generally equipped with water, electricity and gas supplies. However, the exact position in the Beneficiary premises of the each item must be agreed between the suppliers and the Beneficiary prior to the start of deliveries. It is an obligation of the supplier to suggest and formally (i.e. in writing) agree with the Beneficiaries any modification of the water, electricity, gas or other utilities, in case they are necessary, for installation of the equipment according to the technical requirements necessary for its correct installation and performance. Such agreed modifications will be at the expense and responsibility of the supplier."</p> |

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 1